



[FDA Home](#)<sup>3</sup> [Medical Devices](#)<sup>4</sup> [Databases](#)<sup>5</sup>

## Class 2 Device Recall Arrow Glide Thru PeelAway Sheath/Dilator Introducer



[510\(k\)](#)<sup>6</sup> | [De Novo](#)<sup>8</sup> | [Registration & Listing](#)<sup>9</sup> | [Adverse Events](#)<sup>10</sup> | [Recalls](#)<sup>11</sup> | [PMA](#)<sup>12</sup> | [HDE](#)<sup>13</sup> | [Classification](#)<sup>14</sup> | [Standards](#)<sup>15</sup>  
[CFR Title 21](#)<sup>16</sup> | [Radiation-Emitting Products](#)<sup>17</sup> | [X-Ray Assembler](#)<sup>18</sup> | [Medsun Reports](#)<sup>19</sup> | [CLIA](#)<sup>20</sup> | [TPLC](#)<sup>21</sup>

[New Search](#)

[Back to Search Results](#)

### Class 2 Device Recall Arrow Glide Thru PeelAway Sheath/Dilator Introducer



<b>Date Initiated by Firm</b>	June 12, 2017
<b>Date Posted</b>	June 27, 2017
<b>Recall Status</b> <sup>1</sup>	Open <sup>3</sup> , Classified
<b>Recall Number</b>	Z-2699-2017
<b>Recall Event ID</b>	<a href="#">77604</a> <sup>23</sup>
<b>510(K)Number</b>	<a href="#">K122854</a> <sup>24</sup>
<b>Product Classification</b>	Introducer, catheter <sup>25</sup> - <b>Product Code</b> DYB <sup>26</sup>
<b>Product</b>	Arrow Glide Thru Peel-Away Sheath/Dilator Introducer
<b>Code Information</b>	Device Listing # D184260, Material # PL-01055
<b>Recalling Firm/Manufacturer</b>	Arrow International Inc 2400 Bernville Rd Reading PA 19605-9607
<b>For Additional Information Contact</b>	610-378-0131
<b>Manufacturer Reason for Recall</b>	Arrow is recalling additional lots that were identified as part of an active recall. Arrow is recalling the affected product due to the possibility that the catheter peel-away component hub tabs may prematurely detach when the practitioner begins to peel apart the sheath body from the catheter.
<b>FDA Determined Cause</b> <sup>2</sup>	Device Design
<b>Action</b>	Teleflex/Arrow International mailed an Urgent Medical Device Recall Notification Letter to affected customers on 06/12/2017 to inform them of the issue. Arrow requested that customers examine their inventory immediately for the affected lots and discontinue use and quarantine any products with the associated product codes identified in the notice and complete the Recall Acknowledgement Form and fax back to the number included in the notice.
<b>Quantity in Commerce</b>	9,037 units in the U.S. and 4,505 Internationally
<b>Distribution</b>	Distributed to SC, AL, NJ, IN, MA, GA, CA, PA, AZ, VA, WA and Bangkok
<b>Total Product Life Cycle</b>	<a href="#">TPLC Device Report</a> <sup>27</sup>

<sup>1</sup> A record in this database is created when a firm initiates a correction or removal action. The record is updated if the FDA identifies a violation and classifies the action as a recall, and it is updated for a final time when the recall is terminated.

Learn more about [medical device recalls](#)<sup>28</sup>.

<sup>2</sup> Per FDA policy, recall cause determinations are subject to modification up to the point of termination of the recall.

<sup>3</sup> The manufacturer has initiated the recall and not all products have been corrected or removed. This record will be updated as the status changes.