

# **Recall Action Notification**

Safil Sutures

## Important information on the System for Australian Recall Actions

The TGA publishes information about therapeutic goods supplied in the Australian market that have been subject to a recall action in a publicly searchable database.

Recall action means action taken by the responsible entity (being the person who is responsible for taking the recall action) to resolve a problem with therapeutic goods supplied in the Australian market that have, or may potentially have, deficiencies relating to safety, quality, efficacy (performance) or presentation.

- Recall actions include: the permanent removal of therapeutic goods from supply in the market, the
  taking of corrective action in relation to therapeutic goods (such as repair, modification, adjustment or
  relabelling) and, in the case of medical devices that have been implanted into patients, the issuing of a
  hazard alert containing information for health practitioners on how to manage patients.
- More information about Australian recall actions is available at <a href="http://tga.gov.au/safety/recalls-about.htm">http://tga.gov.au/safety/recalls-about.htm</a>
- If you are taking a medicine, using a medical device or have had a medical device implanted into you, that is the subject to a recall action, and you have any concerns you should seek advice from a health professional. <a href="http://www.healthdirect.org.au/">http://www.healthdirect.org.au/</a>

#### About the release of this information

While reasonable care is taken to ensure that the information is an accurate record of recall actions that responsible entities have reported to the TGA or of which the TGA has become aware, the TGA does not guarantee or warrant the accuracy, reliability, completeness or currency of the information or its usefulness in achieving any purpose.

To the fullest extent permitted by law, including but not limited to section 61A of the Therapeutic Goods Act 1989, the TGA will not be liable for any loss, damage, cost or expense incurred in or arising by reason of any person relying on this information.

The information contained in the SARA database is released under s 61(5C) of the Therapeutic Goods Act 1989. Copyright restrictions apply to the System of Australian Recall actions (SARA) <a href="http://tga.gov.au/about/website-copyright.htm">http://tga.gov.au/about/website-copyright.htm</a>.

### Recall detail

Type of Product <sup>i</sup>	Medical Device
TGA Recall Reference <sup>ii</sup>	RC-2018-RN-00737-1
Product Name/Description <sup>iii</sup>	Safil Sutures
	Safil Violet 1 (4) 70cm HR26
	Safil Violet 4/0 (1,5) 70cm HR22
	Safil Violet Undyed 4/0 (1,5) 70cm, DS24
	Safil Violet Undyed 2/0 (3) 70cm HR26
	Safil Violet 3/0 (2) 70cm, HR22
	Safil Violet 2 (5) 70cm HR 37S Safil Violet 2 (5) 70cm HR37S
	Calli Violet 2 (3) 700111111070
	Multiple Item Codes and Batch Numbers
	ARTG Number: 222738
Recall Action Leveliv	Hospital
Recall Action Classification <sup>v</sup>	Class II
Recall Action Commencement Date <sup>vi</sup>	20/06/2018
Responsible Entity <sup>vii</sup>	B Braun Australia Pty Ltd
Reason / Issue <sup>viii</sup>	B. Braun has been identified that some of the items in specific batches have damaged
	packaging which may affect the sterility of the products. No specific monitoring of patients
	is recommended at this time.
	To date, there have been no reports of any adverse patient outcome which could be
Book II And Look	associated with this potential sterility issue.
Recall Actionix	Recall
Recall Action Instructions <sup>x</sup>	B. Braun is requesting customers:
Instructions <sup>*</sup>	I. Identify, quarantine and return affected devices;
	2. Do not use affected devices anymore;
	Patients with affected devices in use should be monitored carefully;
	4. Inform the responsible personnel in the relevant facilities; and
	5. Confirm receipt of the correspondence as per the instructions given in the Customer
	Letter.
Contact Informationxi	1800 251 705 - B Braun

## **Footnotes**

<sup>&</sup>lt;sup>i</sup> Type of Product: Medicine, Medical Device, or Biological

ii TGA Recall Reference: Unique number given by the TGA

- iii Product Name/Description: Brand name (including active ingredient for medicines) and may include generic reference for the kind of medical devices. Includes all necessary information such as affected: catalogue / model and / or batch / serial numbers.
- iv Recall Action Level: The level to which the recall action is to be undertaken. This is based on the significance of the risk and the channels through which the goods have been distributed. The recall action levels are / Wholesale / Hospital / Retail / Consumer.
  - Wholesale includes wholesalers and state purchasing authorities.
  - Hospital includes nursing homes and institutions, hospital pharmacists, ambulance services, blood and tissue banks and laboratories as well as wholesale as appropriate.
  - Retail includes retail pharmacists, medical, dental and other health care professionals as well as wholesale
    and hospital as appropriate.
  - Consumer includes patients and consumers, as well as wholesale, hospital and retail levels as appropriate.
- <sup>v</sup> Recall Action Classification: Recall actions of therapeutic goods are classified based on the potential risk the deficiency poses to patients / consumers. They are classified as Class I, Class II or Class III.
  - Class I recall action occurs when the product deficiency is potentially life-threatening or could cause a serious
    risk to health.
  - Class II recall action occurs when the product deficiency could cause illness, injury or result in mistreatment, but are not Class I.
  - Class III recall action occurs when the product deficiency may not pose a significant hazard to health, but action may be initiated for other reasons eg. quality related issues.
- vi Recall Action Commencement Date: The date the recall strategy and communication was agreed by the TGA.
- vii Responsible Entity: Sponsor / Supplier / Importer responsible for the recall actions.
- viii Reason / Issue: Reason for the recall action.
- <sup>ix</sup> Recall Action: Recall action is an action taken to resolve a problem with a therapeutic good already supplied in the market for which there are issues or deficiencies in relation to safety, quality, efficacy (performance) or presentation.
- There are three distinct recall actions recall, recall for product correction and hazard alert.
  - Recall The permanent removal of an affected therapeutic good from supply or use in the market.
  - Recall for product correction Repair, modification, adjustment or re-labelling of a therapeutic good. The corrective action may take place at the user's premises or any other agreed location.
  - Hazard alert Information issued to healthcare professionals about issues or deficiencies relating to an implanted medical device or biological product and advice about the ongoing management of patients.
- x Recall Action Instructions: What the customer should do.
- xi Contact Information: Who the customer should contact for additional information and clarification regarding the recall action.