

FDA Home<sup>3</sup> Medical Devices<sup>4</sup> Databases<sup>5</sup>

**Class 1 Device Recall Covidien Trellis 6 Peripheral Infusion System**




6 510(k)<sup>7</sup>|DeNovo<sup>8</sup>|Registration & Listing<sup>9</sup>|Adverse Events<sup>10</sup>|Recalls<sup>11</sup>|PMA<sup>12</sup>|Classification<sup>13</sup>|Standards<sup>14</sup>  
 CFR Title | Radiation-Emitting | X-Ray | Medsun | CLIA<sup>19</sup>|TPLC<sup>20</sup>|Inspections<sup>21</sup>  
 21<sup>15</sup> Products<sup>16</sup> Assembler<sup>17</sup> Reports<sup>18</sup>

[New Search](#)

[Back to Search Results](#)

**Class 1 Recall  
 Covidien Trellis 6 Peripheral  
 Infusion System**



<b>Date Posted</b>	February 03, 2015
<b>Recall Status<sup>1</sup></b>	Open
<b>Recall Number</b>	Z-1017-2015
<b>Recall Event ID</b>	<u>70048</u> <sup>23</sup>
<b>Premarket Notification 510(K) Number</b>	<u>K071664</u> <sup>24</sup>
<b>Product Classification</b>	<u>Catheter, Continuous Flush</u> <sup>25</sup> - <b>Product Code</b> <u>KRA</u> <sup>26</sup>
<b>Product</b>	Covidien Trellis-6 Peripheral Infusion System. Models BVT608010V01, BVT608030V01, BVT612010V01, and BVT612030V01. Sterile EO. Product Usage Intended for controlled and selective infusion of physician-specified fluids, including thrombolytics, into the peripheral vasculature.
<b>Code Information</b>	Model BVT608010, Lot numbers: 9853174, 9864497, 9890772, 9890931, 9925676, 9932672, 9937308, 9937315, 9940474, 9940614, 9976340, 9978233, A001589, A001740. Model BVT608030, Lot Numbers: 9853331, 9887597, 9887695, 9887757, 9932042. Model BVT612010, Lot numbers: 9854121, 9868214, 9868311, 9875472, 9875632, 9876174, 9883295, 9884788, 9884951, 9886196, 9886312, 9932096, 9968511, 9974109, 9974126, 9976472, 9977091, 9978862, A002299, A002423. Model BVT312030, Lot numbers: 9854124, 9892081, 9941213, 9971096, 9971139, 9976506, 9977140, 9978964, 9979548, A003021, A003121, A003651, A007900, A008492, A008603, A008687.
<b>Recalling Firm/ Manufacturer</b>	Covidien 4600 Nathan Ln N Plymouth, Minnesota 55442-2890
<b>For Additional Information Contact</b>	Covidien Customer Service 800-716-6700
<b>Manufacturer Reason for Recall</b>	A manufacturing error resulted in the risk of incorrect proximal and distal balloon inflation port identification on the units. Units have been identified to have the distal balloon inflation port incorrectly labeled as proximal, and, the proximal balloon port incorrectly labeled as distal.
<b>FDA Determined Cause<sup>2</sup></b>	COMPONENT CONTROLS (GMP - GOOD MANUFACTURING PRACTICE): Mix-Up of Materials/Components
<b>Action</b>	Consignees were sent an Urgent Product Recall letter dated 12/15/2014. The letter described the issue, identified the affected product, and the required actions. Affected product was to be returned directly to customers' Covidien Sales Rep; or customers were to contact Covidien Service at 1-800-716-6700 to arrange for product return. Questions can be directed to the Covidien Sales Rep or to Covidien Service at 1-800-716-6700. Customers were to complete and return the Verification Form along with the unused product as soon as possible.
<b>Quantity in Commerce</b>	216 devices (207 US, 9 OUS)
<b>Distribution</b>	Worldwide Distribution - US Nationwide and the countries Australia, Canada, Finland, France, Germany, Ireland, Italy, Spain, Sweden, Switzerland, Turkey, and United Kingdom.

<b>for Recall</b>	port identification on the units. Units have been identified to have the distal balloon inflation port incorrectly labeled as proximal, and, the proximal balloon port incorrectly labeled as distal.
<b>FDA Determined Cause <sup>2</sup></b>	COMPONENT CONTROLS (GMP - GOOD MANUFACTURING PRACTICE): Mix-Up of Materials/Components
<b>Action</b>	Consignees were sent an Urgent Product Recall letter dated 12/15/2014. The letter described the issue, identified the affected product, and the required actions. Affected product was to be returned directly to customers' Covidien Sales Rep; or customers were to contact Covidien Service at 1-800-716-6700 to arrange for product return. Questions can be directed to the Covidien Sales Rep or to Covidien Service at 1-800-716-6700. Customers were to complete and return the Verification Form along with the unused product as soon as possible.
<b>Quantity in Commerce</b>	1032 devices (919 US, 113 OUS)
<b>Distribution</b>	Worldwide Distribution - US Nationwide and the countries Australia, Canada, Finland, France, Germany, Ireland, Italy, Spain, Sweden, Switzerland, Turkey, and United Kingdom.
<b>Total Product Life Cycle</b>	<a href="#">TPLC Device Report<sup>27</sup></a>

<sup>1</sup> For details about termination of a recall see [Code of Federal Regulations \(CFR\) Title 21 §7.55<sup>28</sup>](#)

<sup>2</sup> Per FDA policy, recall cause determinations are subject to modification up to the point of termination of the recall.

**510(K) Database**                      [510\(K\)s with Product Code = KRA and Original Applicant = COVIDIEN<sup>29</sup>](#)

---

**Links on this page:**

1. <http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdomain>
2. <http://www.addthis.com/bookmark.php>
3. <http://www.fda.gov/default.htm>
4. <http://www.fda.gov/MedicalDevices/default.htm>
5. <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/Databases/default.htm>
6. </scripts/cdrh/devicesatfda/index.cfm>
7. </scripts/cdrh/cfdocs/cfPMN/pmn.cfm>
8. </scripts/cdrh/cfdocs/cfpmn/denovo.cfm>
9. </scripts/cdrh/cfdocs/cfRL/rl.cfm>
10. </scripts/cdrh/cfdocs/cfMAUDE/TextSearch.cfm>
11. </scripts/cdrh/cfdocs/cfRES/res.cfm>
12. </scripts/cdrh/cfdocs/cfPMA/pma.cfm>
13. </scripts/cdrh/cfdocs/cfPCD/classification.cfm>
14. </scripts/cdrh/cfdocs/cfStandards/search.cfm>
15. </scripts/cdrh/cfdocs/cfCFR/CFRSearch.cfm>
16. [/scripts/cdrh/cfdocs/cfPCD\\_RH/classification.cfm](/scripts/cdrh/cfdocs/cfPCD_RH/classification.cfm)
17. </scripts/cdrh/cfdocs/cfAssem/assembler.cfm>
18. </scripts/cdrh/cfdocs/Medsun/searchReportText.cfm>
19. </scripts/cdrh/cfdocs/cfClia/Search.cfm>
20. </scripts/cdrh/cfdocs/cfTPLC/tplc.cfm>
21. </scripts/cdrh/cfdocs/cfTPLC/inspect.cfm>
22. <http://www.fda.gov/safety/recalls/enforcementreports/default.htm>
23. [/scripts/cdrh/cfdocs/cfRES/res.cfm?start\\_search=1&event\\_id=70048](/scripts/cdrh/cfdocs/cfRES/res.cfm?start_search=1&event_id=70048)
24. </scripts/cdrh/cfdocs/cfPMN/pmn.cfm?ID=K130904>

FDA Home<sup>3</sup> Medical Devices<sup>4</sup> Databases<sup>5</sup>

**Class 1 Device Recall Covidien Trellis 8 Peripheral Infusion System**



510(k)<sup>7</sup> | DeNovo<sup>8</sup> | Registration & Listing<sup>9</sup> | Adverse Events<sup>10</sup> | Recalls<sup>11</sup> | PMA<sup>12</sup> | Classification<sup>13</sup> | Standards<sup>14</sup>  
 CFR Title | Radiation-Emitting | X-Ray | Medsun | CLIA<sup>19</sup> | TPLC<sup>20</sup> | Inspections<sup>21</sup>  
 21<sup>15</sup> | Products<sup>16</sup> | Assembler<sup>17</sup> | Reports<sup>18</sup>

[New Search](#)

[Back to Search Results](#)

**Class 1 Recall  
 Covidien Trellis 8 Peripheral  
 Infusion System**



<b>Date Posted</b>	February 03, 2015
<b>Recall Status<sup>1</sup></b>	Open
<b>Recall Number</b>	Z-1018-2015
<b>Recall Event ID</b>	<u>70048</u> <sup>23</sup>
<b>Premarket Notification 510(K) Number</b>	<u>K130904</u> <sup>24</sup>
<b>Product Classification</b>	<u>Catheter, Continuous Flush</u> <sup>25</sup> - <b>Product Code</b> <u>KRA</u> <sup>26</sup>
<b>Product</b>	Covidien Trellis-8 Peripheral Infusion System. Models CVT808015, CVT808025, CVT812015, and CVT812025. The following models are not offered for sale in the US: EVT808015, EVT808025, EVT812015, and EVT812025 Sterile EO. Product Usage: Intended for controlled and selective infusion of physician-specified fluids, including thrombolytics, into the peripheral vasculature.
<b>Code Information</b>	Model CVT808015, Lot numbers: 9895996, 9904675, 9910270, 9911926, 9912084, 9912200, 9922490, 9922584, 9922597, 9922601, 9922604, 9922606, 9922617, 9922620, 9922624, 9922635, 9923420, 9924840, 9928078, 9932274, 9933341, 9933774, 9933839, 9933934, 9934029, 9934440, 9934633, 9934686, 9936420, 9937544, 9937660, 9942679, 9943032, 9945436, 9948145, 9949409, 9950000, 9950703. Model CVT808025, Lot numbers: 9904237, 9906429, 9910270, 9912003, 9912117, 9912261, 9922644, 9922658, 9922661, 9922663, 9922739, 9922862, 9922868, 9923391, 9923401, 9923405, 9923428, 9923437, 9923454, 9924796, 9924802, 9924826, 9924847, 9924853, 9928334, 9928629, 9929004, 9929248, 9929344, 9933577, 9933797, 9933903, 9934553, 9934772, 9935134, 9935174, 9935241, 9935245, 9935471, 9935523, 9935757, 9935889, 9936140, 9936631, 9936680, 9936702, 9936735, 9937706, 9938500, 9938960, 9939918, 9939985, 9940234, 9940687, 9940916, 9948001, 9948544, 9949603, 9950150, 9950722, 9951391, 9952404, 9952785, 9952844, 9959913. Model CVT812015, Lot Numbers: 9904401, 9904676, 9910934, 9912387, 9922742, 9922793, 9922886, 9922892, 9922902, 9922904, 9922917, 9923224, 9923226, 9923237, 9923268, 9923275, 9923398, 9923446, 9923453, 9924832, 9932594, 9933474, 9933813, 9933924, 9934192, 9934730, 9936820, 9937095, 9937806, 9942701, 9943235, 9945140. Model CVT812025, Lot Numbers: 9910342, 9911492, 9911683, 9911812, 9912503, 9922498, 9922750, 9922753, 9922873, 9922911, 9923246, 9923403, 9923404, 9923409, 9923426, 9923442, 9923450, 9923457, 9924848, 9925351, 9933666, 9934311, 9934575, 9934700, 9935143, 9935199, 9935494, 9935636, 9935900, 9936109, 9936173, 9936570, 9936597, 9936848, 9937896, 9937968, 9938060, 9938219, 9938831, 9938896, 9941711, 9945304, 9947559, 9948088, 9949321, 9949934, 9950324, 9951247, 9951521, 9952927, 9957277, 9957530, 9959854, 9959944, 9960241, 9960324, 9960635, 9960807, 9961583, 9961682, 9962504, 9963558, 9964120, 9964347, 9964448, 9965066, 9965335, 9966329, 9966429, 9966999, 9967656, 9967694, 9968180, 9969396, 9972428, 9972429, 9972819, 9972867, 9972966.
<b>Recalling Firm/ Manufacturer</b>	Covidien 4600 Nathan Ln N Plymouth, Minnesota 55442-2890
<b>For Additional Information Contact</b>	Covidien Customer Service 800-716-6700
<b>Manufacturer Reason</b>	A manufacturing error resulted in the risk of incorrect proximal and distal balloon inflation