



رقم المحفوظات: ٥٨١٢٥

رقم الصادر: ١٣/١/١٤٤٩

بيروت، في:

٢٦ نيسان ٢٠١٢

جانب نقيب المستشفيات الخاصة في لبنان

الموضوع: إشعار بمتابعة جهاز طبي مغروس

Implants, active, defibrillators, Q-TECH Model 2020

الجهاز المعنى بالمتابعة:

- Implants, active, defibrillators, Q-TECH Model 2020
- Trade Mark: Boston Scientific Ltd
- Local Representative:

بناء على التقارير الصادرة عن الوكالة البريطانية

Medicine and Health Care Products Regulatory Agency (UK) MHRA

والتوصية الصادرة عن الشركة المصنعة والتي تشير الى وجود خلل في عمل الصنف المذكور

اعلاه، نرجو منكم تعميم هذه النشرة على جميع المستشفيات المعنية.

مرفق ربطاً:

- التوصية الصادرة عن الشركة المصنعة

يبلغ:

- دائرة البرامج والمشاريع
- المستشفيات الحكومية
- المحفوظات

مدير عام الصحة
د. وليد عمار

Field Safety Notice



Boston Scientific Ltd

Breakspear Park, Breakspear Way
Hemel Hempstead
Hertfordshire
HP2 4TZ

www.bostonscientific-international.com

07 March 2013

Subject: Software Update Required -- Q-TECH™ Model 2020 programmers and SQ-RX® Model 1010 pulse generators

Dear Customer,

Boston Scientific CRM wants to make you aware of the following product advisory regarding Cameron Health / Boston Scientific subcutaneous defibrillators (S-ICD® System).

Further to this advice, we wish to restate that you follow up with your patients as follows:

- For patients whose device has been implanted for less than three months, we recommend that you schedule a follow-up visit **as soon as possible** within the next six weeks.

This software upgrade is available immediately and will be done as soon as possible by your local Boston Scientific CRM representative.

Please see the attached information regarding this advisory and the acknowledgment form for this information. When completed, please return this acknowledgment form via fax to 01442 411816 or Sign/Scan and email to uk-quality@bsci.com on or before **Thursday 21st March 2013** to acknowledge that you have received, read and understood the information contained and the actions that are required.

If you have any questions, please do not hesitate to contact your local Boston Scientific CRM representative.

Sincerely,

Terry Nopper
UK Sales and Marketing Director
Boston Scientific Ltd.

