Recall detail

Type of Product TGA Recall Reference RC-2014-RN-01366-1 Product Name/Description* Cheek Retractor, f/MatrixMANDIBLE U-shaped, flexible Part Number: 397.232 Multiple lot numbers affected ARTG Number: 153950 Recall Action Class II Recall Action Class II Recall Action Commencement Date Synthes Australia Pty Ltd The Cheek Retractor for MatrixMANDIBLE U-shaped may not function as intended due the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the patent. However, in the event that the spring or a subcomponent of the spring was to fail and/o corrode, the theoretical possibility exists that the particles may transfer from the in trum to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgid delay may also arise if the retractor does not work during a procedure and a replacement of other procedure and a replacement of the spring was to fail and/or corrode, the theoretical possibility exists that the particles may transfer from the in trum to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgid delay may also arise if the retractor does not work during a procedure and a replacement of the spring was to fail and/or corrode, the theoretical possibility exists that the particles may transfer from the in trum to the patient, potentially leading to an adverse events related to this issue. Recall Action Recall Action Recall Action To date, there have been no reports of adverse events related to this issue. Recall Action Recall Action Recall Action Instructions Recall Action Reca			
Cheek Retractor, f/MatrixMANDIBLE U-shaped, flexible Part Number: 397.232 Multiple lot numbers affected ARTG Number: 153950 Recall Action Level® Resol Action Class II Recall Action Class II Reason / Issue® The Cheek Retractor for MatrixMANDIBLE U-shaped may not function as intended due the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the patent. However, in the event that the spring or a subcomponent of the spring was to fail and/or corrode, the theoretical possibility exists that the particles may transfer from the in trum to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgic delay may also arise if the retractor does not work during a procedure and a replacement needs to be found. To date, there have been no reports of adverse events related to this issue. Recall Action® Recall Action Customers are asked to inspect invetory and to follow the instructions provided for any affected units. To mitigate risk, should customers with affected instruments require instrument for emergency cases, please perform a visual and functional inspection, as detailed in the cistomer letter, prior to, during and after use. Customers are asked to exercise caution regarding the risk of corrosion or spring breaking during the procedure.	Type of Producti	Medical Device	
Part Number: 397.232 Multiple lot numbers affected ARTG Number: 153950 Recall Action Level** Hospital Recall Action Class II Class II Synthes Australia Pty Ltd The Cheek Retractor for MatrixMANDIBLE U-shaped may not function as intended due the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the patient. However, in the event that the spring or a subcomponent of the spring was to fail and/or corrode, the theoretical possibility exists that the particles may transfer from the in trum to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgic delay may also arise if the retractor does not work during a procedure and a replacement needs to be found. To date, there have been no reports of adverse events related to this issue. Recall Action Recall Action Customers are asked to inspect invetory and to follow the instructions provided for affected units. To mitigate risk, should customers with affected instruments require the instrument for emergency cases, please perform a visual and functional inspection, as detailed in the cistomer letter, prior to, during and after use. Customers are asked exercise caution regarding the risk of corrosion or spring breaking during the procedure.	TGA Recall Reference	RC-2014-RN-01366-1	
Multiple lot numbers affected ARTG Number: 153950 Recall Action Level ^[v] Hospital Recall Action Classification ^v Class II Recall Action Commencement Date ^{vii} Synthes Australia Pty Ltd Responsible Entity ^{vii} Synthes Australia Pty Ltd The Cheek Retractor for MatrixMANDIBLE U-shaped may not function as intended the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the patent. However, in the event that the spring or a subcomponent of the spring was to fail and/o corrode, the theoretical possibility exists that the particles may transfer from the instrumt to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgid delay may also arise if the retractor does not work during a procedure and a replacement needs to be found. To date, there have been no reports of adverse events related to this issue. Recall Action Recall Action Customers are asked to inspect invetory and to follow the instructions provided for any affected units. To mitigate risk, should customers with affected instruments require the instrument for emergency cases, please perform a visual and functional inspectior, as detailed in the cistomer letter, prior to, during and after use. Customers are asked to exercise caution regarding the risk of corrosion or spring breaking during the procedure.	Product Name/Description ⁱⁱⁱ	Cheek Retractor, f/MatrixMANDIBLE U-shaped, flexible	
Recall Action Level Hospital Recall Action Class II Recall Action Commencement Date Synthes Australia Pty Ltd Responsible Entity Synthes Australia Pty Ltd The Cheek Retractor for MatrixMANDIBLE U-shaped may not function as intended due the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the patent. However, in the event that the spring or a subcomponent of the spring was to fail and/o corrode, the theoretical possibility exists that the particles may transfer from the instrumt to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgic delay may also arise if the retractor does not work during a procedure and a replacement needs to be found. To date, there have been no reports of adverse events related to this issue. Recall Action Recall Action Recall Action Customers are asked to inspect invetory and to follow the instructions provided fo any affected units. To mitigate risk, should customers with affected instruments require the instrument for emergency cases, please perform a visual and functional inspection, as detailed in the cistomer letter, prior to, during and after use. Customers are asked to exercise caution regarding the risk of corrosion or spring breaking during the procedure.		Part Number: 397.232	
Recall Action Classification Classification Classification Commencement Date 24/12/2014 Responsible Entity 3ynthes Australia Pty Ltd The Cheek Retractor for MatrixMANDIBLE U-shaped may not function as intended the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the patent. However, in the event that the spring or a subcomponent of the spring was to fail and/or corrode, the theoretical possibility exists that the particles may transfer from the instrumt to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgic delay may also arise if the retractor does not work during a procedure and a replacement needs to be found. To date, there have been no reports of adverse events related to this issue. Recall Action Recall Action Customers are asked to inspect invetory and to follow the instructions provided for any affected units. To mitigate risk, should customers with affected instruments require the instrument for emergency cases, please perform a visual and functional inspection, as detailed in the cistomer letter, prior to, during and after use. Customers are asked to exercise caution regarding the risk of corrosion or spring breaking during the procedure.		Multiple lot numbers affected	
Recall Action Class II Responsible Entity*ii Synthes Australia Pty Ltd The Cheek Retractor for MatrixMANDIBLE U-shaped may not function as intended due the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the patent. However, in the event that the spring or a subcomponent of the spring was to fail and/o corrode, the theoretical possibility exists that the particles may transfer from the instrumt to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgid delay may also arise if the retractor does not work during a procedure and a replacement needs to be found. To date, there have been no reports of adverse events related to this issue. Recall Action Instructions* Customers are asked to inspect invetory and to follow the instructions provided for any affected units. To mitigate risk, should customers with affected instruments require the instrument for emergency cases, please perform a visual and functional inspection, as detailed in the cistomer letter, prior to, during and after use. Customers are asked to exercise caution regarding the risk of corrosion or spring breaking during the procedure.		ARTG Number: 153950	
Recall Action Commencement Datevi Responsible Entityvii Synthes Australia Pty Ltd The Cheek Retractor for MatrixMANDIBLE U-shaped may not function as intended due the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the patient. However, in the event that the spring or a subcomponent of the spring was to fail and/or corrode, the theoretical possibility exists that the particles may transfer from the instrumt to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgict delay may also arise if the retractor does not work during a procedure and a replacement needs to be found. To date, there have been no reports of adverse events related to this issue. Recall Action Instructions Recall Action Instructions Customers are asked to inspect invetory and to follow the instructions provided for any affected units. To mitigate risk, should customers with affected instruments require the instrument for emergency cases, please perform a visual and functional inspection, as detailed in the cistomer letter, prior to, during and after use. Customers are asked to exercise caution regarding the risk of corrosion or spring breaking during the procedure.	Recall Action Leveliv	Hospital	
Responsible Entity*** Synthes Australia Pty Ltd The Cheek Retractor for MatrixMANDIBLE U-shaped may not function as intended due the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the patient. However, in the event that the spring or a subcomponent of the spring was to fail and/o corrode, the theoretical possibility exists that the particles may transfer from the instrumt to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgid delay may also arise if the retractor does not work during a procedure and a replacement needs to be found. To date, there have been no reports of adverse events related to this issue. Recall Action Instructions* Customers are asked to inspect invetory and to follow the instructions provided for any affected units. To mitigate risk, should customers with affected instruments require the instrument for emergency cases, please perform a visual and functional inspection, as detailed in the cistomer letter, prior to, during and after use. Customers are asked to exercise caution regarding the risk of corrosion or spring breaking during the procedure.		Class II	
The Cheek Retractor for MatrixMANDIBLE U-shaped may not function as intended the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the patent. However, in the event that the spring or a subcomponent of the spring was to fail and/or corrode, the theoretical possibility exists that the particles may transfer from the instrument to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgice delay may also arise if the retractor does not work during a procedure and a replacement needs to be found. To date, there have been no reports of adverse events related to this issue. Recall Action Recall Action Recall Action Customers are asked to inspect invetory and to follow the instructions provided for any affected units. To mitigate risk, should customers with affected instruments require the instrument for emergency cases, please perform a visual and functional inspection, as detailed in the cistomer letter, prior to, during and after use. Customers are asked to exercise caution regarding the risk of corrosion or spring breaking during the procedure.	Recall Action Commencement Datevi	24/12/2014	
the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the patient. However, in the event that the spring or a subcomponent of the spring was to fail and/o corrode, the theoretical possibility exists that the particles may transfer from the instrumt to the patient, potentially leading to an adverse tissue reaction and/or infection. A surgice delay may also arise if the retractor does not work during a procedure and a replacement needs to be found. To date, there have been no reports of adverse events related to this issue. Recall Action Recall Action Customers are asked to inspect invetory and to follow the instructions provided for any affected units. To mitigate risk, should customers with affected instruments require the instrument for emergency cases, please perform a visual and functional inspection, as detailed in the cistomer letter, prior to, during and after use. Customers are asked to exercise caution regarding the risk of corrosion or spring breaking during the procedure.	Responsible Entity ^{vii}	Synthes Australia Pty Ltd	
Customers are asked to inspect invetory and to follow the instructions provided for any affected units. To mitigate risk, should customers with affected instruments require the instrument for emergency cases, please perform a visual and functional inspection, as detailed in the cistomer letter, prior to, during and after use. Customers are asked to exercise caution regarding the risk of corrosion or spring breaking during the procedure.	Reason / Issue ^{viii}	the potential for failure and/or corrosion of the internal spring which has been manufactured from an incorrect material. As an internal part of the instrument, the spring is not in direct contact with the pati However, in the event that the spring or a subcomponent of the spring was to fail a corrode, the theoretical possibility exists that the particles may transfer from the into the patient, potentially leading to an adverse tissue reaction and/or infection. A sidelay may also arise if the retractor does not work during a procedure and a replacement of the pound.	ent. Ind/or trument urgical
affected units. To mitigate risk, should customers with affected instruments require the instrument for emergency cases, please perform a visual and functional inspection, as detailed in the cistomer letter, prior to, during and after use. Customers are asked to exercise caution regarding the risk of corrosion or spring breaking during the procedure.	Recall Actionix	Recall	
Contact Information ^{xi} 1800 796 8437 - Synthes Australia Pty Ltd		affected units. To mitigate risk, should customers with affected instruments require instrument for emergency cases, please perform a visual and functional inspection detailed in the cistomer letter, prior to, during and after use. Customers are asked	the , as to
	Contact Information ^{xi}	1800 796 8437 - Synthes Australia Pty Ltd	

Footnotes

- ⁱ Type of Product: Medicine, Medical Device, or Biological
- ii TGA Recall Reference: Unique number given by the TGA
- iii Product Name/Description: Brand name (including active ingredient for medicines) and may include generic reference for the kind of medical devices. Includes all necessary information such as affected: catalogue / model and / dr batch /