FDA Home³ Medical Devices⁴ Databases⁵

Class 2 Device Recall Lumenis

6 510(k)⁷|DeNovo⁸|Registration & Listing⁹|Adverse Events¹⁰|Recalls¹¹|PMA¹²|Classification¹³|Standards¹⁴

CORM

CFR Title 2115

Radiation-Emitting Products¹⁶

X-Ray Assembler¹⁷ Medsun

|CLIA¹⁹|TPLC²⁰|Inspections²¹

Reports¹⁸

New Search

Back to Search Results

Class 2 Recall Lumenis



Date Posted

SurgerSearch

September 30, 2014

Recall Status¹

Open

Recall Number

Z-2733-2014

Recall Event ID

6913923

Premarket Notification 510(K) Number

K13019524

Product Classification

Laser, Ophthalmic²⁵ - Product Code HQF²⁶

Product

Array LaserLink, Manufactured by Lumenis, The Array" LaserLink" is a laser system accessory intended for use in the treatment of ocular pathology. " For the Posterior Segment, the Array" LaserLink" is indicated for use in Retinal Photocoagulation and Panretinal Photocoagulation of Vascular and Structural Abnormalities of the Retina and Choroid including: "Proliferative and Severe and Very Severe Non-Proliferative Diabetic Retinopathy "Macular Edema associated with Proliferative or Non-Proliferative Diabetic Retinopathy " Choroidal Neovascularization " Retinal Neovascularization associated with Retinal Occlusive Disease (Branch Retinal Vein Occlusion; Central Retinal Vein Occlusion) "Macular Edema associated with Branch Retinal Vein Occlusion "Retinal Tears and Detachments And Anterior Segments as follows: "Iridotomy in Closed Angle Glaucoma" Trabeculoplasty in Open Angle Glaucoma

Code Information

All assembled units since product release, Part Number: GA-0006700.

Recalling Firm/ Manufacturer

Lumenis, Inc. 3959 W 1820 S

Salt Lake City, Utah 84104

For Additional Information Contact Mr. Rick Gavkowski 801-656-2690

Manufacturer Reason

for Recall

Lumenis initiated a field-correction for the Array Laser Link", GA-0006700 (SN XXYYZZ) because the system may project an erroneous pattern display on the retina, which is

different than the desired pattern.

FDA Determined Cause 2

DESIGN: Software Design

Action

Lumenis sent a Safety Advisory Notice dated July 15, 2014, to all affected consignees. The letter identified the product, the problem, and the action to be taken by the consignee. Consignees were instructed that Lumenis would contact consignees to schedule a product update, to be completed at a convenient time. For questions consignees should call 801-656-2549. For questions regarding this recall call 801-656-2690.

Quantity in Commerce

Distribution

Worldwide Distribution - USA (nationwide) and Internationally to JAPAN, CHINA, and CANADA.

Total Product Life Cycle

TPLC Device Report²⁷